



TRUCK SCALE TEST RECORD: 668261

134 Mendon Street, Bellingham, MA. 02019 Phone: (800) 232-0252

Test Procedure Used: WI-50718 in accordance with NIST H44

FINAL TEST In tolerance: Yes ✓

CITY OF BRISTOL TFR STA - PlaceID: 9523210002 Task#: 1886326

ATTN JOSH COREY, BRISTOL, CT 06010

Contact: JOSH COREY Ph: 860-584-7791

Scale Location: **Truck Scale - inbound**

Report Date: 12/10/2025

Inspection Date: 12/10/2025

Last Inspection: On File

Next Inspection Due: 06/2026

Nominal Capacity Sticker: Yes

Test Equipment Used

Test equipment type: NIST Traceable Weights Weight Expiration: On File Weight Used: 25000
 Certificate Number: On File Truck No: 1411 Units of Weights Used: Lb
 Test Equipment Used: NIST Traceable Certified Test Weights

Equipment Specifications

Indicator:	Manufacturer	Model	Serial#	Capacity	(d) div	# d	Deck	Size
Receiver:	Fairbanks Scales	FB6011	200840080033	100000	10	10000	Concrete	60x10
	Fairbanks Scales	plt6010HV408	6330R					

Certified Weights Applied: 25000 Lbs. NIST H44 Tolerance Used on this Test: Maintenance

SHIFT TEST RESULTS	AS FOUND	Sec. 1	Sec. 2	Sec. 3	Sec. 4	Sec. 5	Sec. 6	Sec. 7	Sec. 8	Sec. 9	Sec. 10
		Indicated:	24990	25000	25010	25010					
	Error:	-10	0	10	10						
	Tolerance +/-:	50.0	50.0	50.0	50.0						
AS LEFT	Indicated:	24990	25000	25010	25010						
	Error:	-10	0	10	10						
	Tolerance +/-:	50.0	50.0	50.0	50.0						

LOAD TEST	AS FOUND				AS LEFT			
	Applied	Indicated	Error	Tolerance	Applied	Indicated	Error	Tolerance
0 ZERO:	0				0			
1								
rtn to zero:	0				0			

STRAIN TEST	Test/Tare Load	Test Weights	Total Combined	Instrument Reading	Error
		33110	25000	58110	58120


Actions taken or comments:
Inspected scale, tested scale and scale was good

Condition of scale or serviceable items:
Scale is in good working order. Recommend getting an annual lubrication of loadcells setup.

Inspector

Owner/User


 Anthony Porzio Lic#:





TRUCK SCALE TEST RECORD: 668262

134 Mendon Street, Bellingham, MA. 02019 Phone: (800) 232-0252

Test Procedure Used: WI-50718 in accordance with NIST H44

FINAL TEST In tolerance: Yes ✓

CITY OF BRISTOL TFR STA - PlaceID: 9523210002 Task#: 1886326

ATTN JOSH COREY, BRISTOL, CT 06010

Contact: JOSH COREY Ph: 860-584-7791

Scale Location: **Truck Scale - Outbound**

Report Date: 12/10/2025

Inspection Date: 12/10/2025

Last Inspection: On File

Next Inspection Due: 06/2026

Nominal Capacity Sticker: Yes

Test Equipment Used

Test equipment type:	NIST Traceable Weights	Weight Expiration:	On File	Weight Used:	25000
Certificate Number:	On File	Truck No:	1411	Units of Weights Used:	Lb
Test Equipment Used:	NIST Traceable Certified Test Weights				

Equipment Specifications

Indicator:	Manufacturer	Model	Serial#	Capacity	(d) div	# d	Deck	Size
Receiver:	Fairbanks Scales	FB6011	180580140064	100000	10	10000	Concrete	60x10
	Fairbanks Scales	plt6010HV408	6331R					

Certified Weights Applied: 25000 Lbs. NIST H44 Tolerance Used on this Test: Maintenance

SHIFT TEST RESULTS	AS FOUND	Sec. 1	Sec. 2	Sec. 3	Sec. 4	Sec. 5	Sec. 6	Sec. 7	Sec. 8	Sec. 9	Sec. 10
		Indicated:	24990	24990	25000	25000					
Error:	-10	-10	0	0							
Tolerance +/-:	50.0	50.0	50.0	50.0							
AS LEFT	Indicated:	24990	24990	25000	25000						
	Error:	-10	-10	0	0						
Tolerance +/-:	50.0	50.0	50.0	50.0							

LOAD TEST	AS FOUND				AS LEFT			
	Applied	Indicated	Error	Tolerance	Applied	Indicated	Error	Tolerance
0 ZERO:	0				0			
1								
rtn to zero:	0				0			

STRAIN TEST	Test/Tare Load	Test Weights	Total Combined	Instrument Reading	Error
		33110	25000	58110	58100

Actions taken or comments:
Inspected scale, tested scale and scale was good

Condition of scale or serviceable items:
Scale is in good working order. Recommend getting an annual lubrication of loadcells setup.

Inspector

Owner/User


Anthony Porzio Lic#:

